



Robert Wood Johnson
Medical School

**Stat Pak & Uni-Gold
TEMPERATURE LOG**

New Jersey Rapid HIV Testing Support Program
Location
Somerset, NJ 08873
Fax: 732 235-9012

SITE NAME:
SITE:#

MONTH/YEAR:

DATE	TEST KITS				KIT CONTROLS				Operator
	StatPak		Unigold		StatPak		Unigold		
	8°-30° C (46°-86° F) TEST KITS (Min/Max)		2°-27° C (36°-81° F) TEST KITS (Min/Max)		2°-8° C (36°-46° F) CONTROLS (Min/Max)		2°-8° C (36°-46° F) CONTROLS (Min/Max)		
	Min	Max	Min	Max	Min	Max	Min	Max	
1.	/		/		/		/		
2.	/		/		/		/		
3.	/		/		/		/		
4.	/		/		/		/		
5.	/		/		/		/		
6.	/		/		/		/		
7.	/		/		/		/		
8.	/		/		/		/		
9.	/		/		/		/		
10.	/		/		/		/		
11.	/		/		/		/		
12.	/		/		/		/		
13.	/		/		/		/		
14.	/		/		/		/		
15.	/		/		/		/		
16.	/		/		/		/		
17.	/		/		/		/		
18.	/		/		/		/		
19.	/		/		/		/		
20.	/		/		/		/		
21.	/		/		/		/		
22.	/		/		/		/		
23.	/		/		/		/		
24.	/		/		/		/		
25.	/		/		/		/		
26.	/		/		/		/		
27.	/		/		/		/		
28.	/		/		/		/		
29.	/		/		/		/		
30.	/		/		/		/		
31.	/		/		/		/		

Please use a new sheet each month.
Please record the minimum → maximum temperatures recorded each day
SECTION 2.7.4